



21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

Article-No.: **14-100**

Description: Black 300 Latex

Latex examination glove black, non sterile, powder free

THICKNESS		
cuff	palm	fingertips
0,12 mm	0,16 mm	0,18 mm



PRODUCT DESCRI	PTION						
material	✓ Latex	Nitrile	☐ Vinyl	☐ Vinyl-Nitrile-	☐ Polyethy-lene	☐ TPE	cotton
				mixture	(PE)		
colour	☐ white	☐ blue	☑ black	☐ mint	☐ purple	☐ mix	apple-green
characteristics	☐ prepowdered	powderfree	☐ sterile	non sterile	☑ ambidex-	☐ fits hand-	☐ biodgra-
					trous	specific	dable
surface	✓ full textured	□ finger	☐ not textured	embossed		side	
		textured					
01750							
SIZES	VC (F. C)	S / S 7 \	NA (7. 0)	L (8-9)	VI (0.40)	VVI (40 44)	VVVI (44 42)
width	XS (5-6) ≤ 80 mm	S (6-7) 80 ± 10 mm	M (7-8) 95 ± 10 mm	110 ± 10 mm	XL (9-10) 115 ± 10 mm	XXL (10-11)	XXXL (11-12)
length	≥ 300 mm	≥ 300 mm	2 300 mm	≥ 300 mm	≥ 300 mm	_	_
		2 300 IIIII	2 300 Hilli	2 300 mm	2 300 IIIIII	_	_
REGULATORY AFF	AIRS						
PPE-Regulation	☐ Category I	☐ Category II	Category III	☐ no PPE-article			
(EU) 2016/425							CE
MD-Regulation	✓ Class I	☐ Class II a	☐ Class III		☐ measuring	☐ no medical	6
(EU) 2017/745					function	device	
Food Contact		☑ aqueous	☐ fatty foods	alcoholic	☐ dry foods	☐ not approved	ΩÏ
(EG) 1935/2004		foods		foods		for food-	201
						contact	
STANDARDISATIO	N						
EN 388 Mechanical	abrasion	blade cut	tear resistance	puncture	blade cut	impact test	
Risks	resistance	resistance		resistance	resistance		
		Coupe-Test			TDM-Test		
Level	not applicable						
EN 374-1	chemical		code letter	level	permeation time	degradation	
Chemical Risks	Sodium hydroxide	40%	K	6	> 480 min	-19,2 %	ISO 374-1/Type B
	Hydrogen Peroxide		Р	3	> 60 min	-14,0 %	
EN 374-4			T	2	> 30 min	-66,3 %	()(-)
	Formaldehyde 37%	6	'		> 30 IIIIII	· · · · · · · · · · · · · · · · · · ·	1 \ \(\D \)
Degradation	Formaldehyde 37%	6	ı	2	> 30 IIIII		
Degradation	Formaldehyde 37%	6	I I	2	> 30 mm	,	WAT.
Degradation							KPT
EN 374-5				and fungi). Test acco			EN 150 374-5:2016
EN 374-5 microorganism							
EN 374-5 microorganism tightness	The glove is tight a	gainst microorganis	ms (viral, bacteria a	and fungi). Test acco			EN ISO 374-5:2016
EN 374-5 microorganism tightness EN ISO 21420	The glove is tight a		ms (viral, bacteria a	and fungi). Test acco			EN ISO 374-5:2016
EN 374-5 microorganism	The glove is tight a	gainst microorganis	ms (viral, bacteria a	and fungi). Test acco			EN 150 374-5:2016
EN 374-5 microorganism tightness EN ISO 21420 protective gloves	The glove is tight a	gainst microorganis ne requirements acc	ms (viral, bacteria a	and fungi). Test acco	ording to ISO 16604		EN ISO 374-52016 VIRUS
EN 374-5 microorganism tightness EN ISO 21420 protective gloves	The glove is tight a	gainst microorganis ne requirements acc	ms (viral, bacteria a	and fungi). Test acco	ording to ISO 16604		EN IND 374-52016 VIRUS
EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for	The glove is tight a	gainst microorganis ne requirements acc	ms (viral, bacteria a	and fungi). Test acco	ording to ISO 16604		EN ISO 374-52016 VIRUS
EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for	The glove is tight a The glove meets th	gainst microorganis ne requirements acc ne requirements acc	cording to EN ISO 21	and fungi). Test acco	ording to ISO 16604	- method B.	EN INO 374-3-2016 VIRUS
EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	The glove is tight a The glove meets th	gainst microorganis ne requirements acc ne requirements acc	cording to EN ISO 21	and fungi). Test acco	ording to ISO 16604	- method B.	EN ISO 378-3:2016 VIRUS EN 455 AQL
EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	The glove is tight a The glove meets th The glove meets th	gainst microorganis ne requirements acc ne requirements acc	cording to EN ISO 21	and fungi). Test acco	ording to ISO 16604	- method B.	EN ISO 374-52016 VIRUS EN 455
EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use	The glove is tight a The glove meets th The glove meets th	gainst microorganis ne requirements acc ne requirements acc	cording to EN ISO 21	and fungi). Test acco	ording to ISO 16604	- method B.	EN ISO 374-52016 VIRUS EN 455 AQL
EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	The glove is tight a The glove meets th The glove meets th The glove has an A general Inspection	gainst microorganis ne requirements acc ne requirements acc	cording to EN ISO 21	and fungi). Test acco	ording to ISO 16604	- method B.	EN ISO 374-52016 VIRUS EN 455 AQL



21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de

omepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

Article-No.: **14-100**

Description: Black 300 Latex

Latex examination glove


black, non sterile, powder free

LOGISTIC DATA S	UBPACKING		
generell informatior	1		
material		carton	
pieces per subpacking		100	
GTIN subpacking size XS		4044941729923	
GTIN subpacking size S		4044941729947	
GTIN subpacking size M		4044941729961	
GTIN subpacking size L		4044941729985	
GTIN subpacking size XL		4044941730004	
GTIN subpacking size XXL		=	
GTIN subpacking size XXXL		-	
PZN subpacking size XS		-	
PZN subpacking size S		-	
PZN subpacking size M		-	
PZN subpacking size	L	=	
PZN subpacking size XL		-	
PZN subpacking size XXL		-	
PZN subpacking size XXXL		-	
measures & size			
length		280 mm	
width		127 mm	
heigth		75 mm	
weights			
size	net weight	gross weight	
XS	700 g	760 g	
S	760 g	820 g	
М	830 g	890 g	
L	850 g	910 g	
XL	940 g	1.000 g	
XXL	-	-	
XXXL	-	-	

LOGISTIC DATA	PALETTE	
general informatio	n	
kind of palett		euro-palette
measures & size		
cartons per layer		8
layers per palette		6
heigth of the palett	e	192 cm
weights		
size	net weight	gross weight
XS	389 g	414 g
S	418 g	443 g
M	451 g	476 g
L	461 g	486 g
XL	504 g	529 g
XXL	=	-
XXXL	-	-



	OUTER PACKING		
generell information	on		
material		carton	
subpackings per outer packing		10	
GTIN outer packing size XS		4044941729930	
GTIN outer packing size S		4044941729954	
GTIN outer packing size M		4044941729978	
GTIN outer packing size L		4044941729992	
GTIN outer packing size XL		4044941730011	
GTIN outer packing size XXL		-	
GTIN outer packing size XXXL		-	
PZN outer packing:	size XS	-	
PZN outer packing:	size S	-	
PZN outer packing size M		-	
PZN outer packing size L		-	
PZN outer packing size XL		-	
PZN outer packing size XXL		-	
PZN outer packing size XXXL		=	
measures & size			
length		390 mm	
width		270 mm	
heigth		295 mm	
weights			
size	net weight	gross weight	
XS	7.600 g	8.100 g	
S	8.200 g	8.700 g	
M	8.900 g	9.400 g	
L	9.100 g	9.600 g	
XL	10.000 g	10.500 g	
XXL	-	-	
XXXL -		-	





21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

Article-No.: 14-100

Description: Black 300 Latex

Latex examination glove

black, non sterile, powder free



WARNINGS AND SAFETY INFORMATION

storage / expiry date

Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging.

use and control

Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical.

Before use, the gloves must be checked for holes or damage.

disposal

Used gloves must be disposed of after contact with chemicals in accordance with the disposal regulations for the chemical and the regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.

disinfection

Disinfection is not intended for these gloves and is the responsibility of the user.

warnings/ allergy information Protective gloves are intended for single use only.

This product contains dithiocarbamates and natural latex, which can trigger allergic reactions, including anaphylactic reactions

donning and doffing instructions











rev-no.:

date 03. 09.2024

changes and errors excepted